


CALIBRATE THERMOMETERS OFTEN!

Make sure temperatures are accurate

Ice water method



Wait 30 seconds

32°F

LOG


Date: _____
 Time: _____
 How used: _____
 Temperature: _____
 Action taken: _____
 Initials: _____

Adjust



Hold 30 seconds
 Recheck

Boiling water method



Wait 30 seconds

212°F

What to do

- ⊕ Fill a large glass with crushed ice.
- ⊕ Add clean tap water until glass is full.
- ⊕ Put the thermometer or probe stem into the ice water so that the sensing area is completely submerged.
- ⊕ Wait 30 seconds. 32°F
- ⊕ Hold the adjusting nut securely with a wrench or pliers and rotate the head of the thermometer until it reads 32°F.

Notes

- ⊕ Stir the mixture well
- ⊕ Do not let the stem touch the bottom or sides of the glass.
- ⊕ The thermometer stem or probe stem must remain in the ice water.
- ⊕ Press the reset button on a digital thermometer to adjust the readout.

What to do

- ⊕ Bring clean tap water to a boil in a deep pan.
- ⊕ Put the thermometer or probe stem into the boiling water so that the sensing area is completely submerged.
- ⊕ Wait 30 seconds. 212°F
- ⊕ Hold the adjusting nut securely with a wrench or pliers and rotate the head of the thermometer until it reads 212°F (100°C) or the appropriate boiling temperature.

Notes

- ⊕ Do not let the stem touch the bottom or sides of the pan.
- ⊕ The thermometer stem or probe stem must remain in the boiling water.
- ⊕ Press the reset button on a digital thermometer to adjust the readout.



WOOD COUNTY
HEALTH DISTRICT
 1840 E. Gypsy Lane Rd.
 Bowling Green, OH 43402
 (419) 354-2702